

# NITRILE GLOVES

100 PER BOX



## APPENDIX A: PACKAGING



## APPENDIX B: PHYSICAL DIMENSIONS

### Gloves Physical Dimensions

Dimensions	Standards		
	Gloves	ASTM D3578	EN 455
Length (mm)	Min 230 Min 240 300 ± 10	Min 220 ( XS , S) Min 230 ( M , L , XL)	Min 240
Palm width (mm)			
• XS	76 ± 3	70 ± 10	≤80
• S	84 ± 3	80 ± 10	80 ± 10
• M	94 ± 3	95 ± 10	95 ± 10
• L	105 ± 3	110 ± 10	110 ± 10
• XL	113 ± 3	120 ± 10	≥110
Thickness:			
Single wall (mm)	0.09±0.02	± 0.07	N/A
• Finger	0.07±0.02	± 0.07	N/A
• Palm			N/A
Properties	ASTM D6319	EN 455	
Tensile Strength (MPa)			
• Before Aging	Min 14	N/A	
• After Aging	Min 14	N/A	
Elongation at Break (%)			
• Before Aging	Min 500	N/A	
• After Aging	Min 400	N/A	
Median Force at Break (N)			
• Before Aging	N/A	Min 6	
• After Aging	N/A	Min 6	

**TEST REPORT:**

Date: 09 JAN 2014 Tel: +65 68851312 Fax: +65 67784301  
 Client's Ref: 221403179Rev1 Email: zhou.xiao@tuv-sud-psb.sg

**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



PSB Singapore

Choose certainty.  
Add value.

**SUBJECT**

Overall Migration Test for "Powder Free Nitrile Exam Gloves" Sample

**CLIENT****SAMPLE SUBMISSION DATE**

16 Oct 2013 &amp; 30 Oct 2013

**DESCRIPTION OF SAMPLE**

A packet of sample labelled as follows was received. The tests were confirmed to be analysed on 13 Dec 2013.

Glove Type	Code	Lot No.	Size
Powder Free Nitrile Exam Gloves	N28FFB	TUV141013-01	M

**DATE OF ANALYSIS**

18 Dec 2013 – 06 Jan 2014

**METHOD OF TEST**

The sample was analysed for the following tests according to Commission Regulation (EU) No 10/2011.

- Preparation of Test Specimen**  
Only the exterior of the glove sample was performed for the test.
- Overall Migration Content with Aqueous Food Simulant (10% Ethanol, 3% Acetic Acid, 20% Ethanol & 50% Ethanol)**  
According to BS EN 1186-9:2002 – Test Methods for overall migration into aqueous food simulants by article filling.
- Overall Migration Content with Fatty Food Simulant (Olive Oil)**  
According to BS EN 1186-8:2002 – Test Methods for overall migration into olive oil by article filling.



Laboratory:  
TUV SUD PSB Pte. Ltd.  
No. 1 Science Park Drive  
Singapore 118221

Phone : +65-6885 1333  
Fax : +65-6776 8670  
E-mail: testing@tuv-sud-psb.sg  
www.tuv-sud-psb.sg  
Co. Reg : 19902667R

Regional Head Office:  
TUV SUD Asia Pacific Pte. Ltd.  
3 Science Park Drive, #04-01/05  
The Franklin, Singapore 118223  
TUV®

**TEST REPORT:**

09 JAN 2014



PSB Singapore

**RESULTS**

Table 1 : Overall Migration Content with Food Simulants for the "Powder Free Nitrile Exam Gloves, N28FFB, Lot No.:TUV141013-01" Sample

Type of Simulant	Testing Condition	Surface Area (dm <sup>2</sup> )	Volume of Extractant (ml)	Overall Migration (mg/dm <sup>2</sup> )	Commission Regulation (EU) No 10/2011 Requirement for Overall Migration Content (mg/dm <sup>2</sup> )
1. 10% Ethanol	40 °C, 2 hours	5.14	250	<1.0	<10
2. 3% Acetic Acid	40 °C, 2 hours	5.08	250	1.6	<10
3. 20% Ethanol	40 °C, 2 hours	5.11	250	<1.0	<10
4. 50% Ethanol	40 °C, 2 hours	5.12	250	2.6	<10
5. Vegetable Oil (Olive Oil)	40 °C, 2 hours	5.06	250	<1.0	<10

Based on the above results, the "Powder Free Nitrile Exam Gloves, N28FFB, Lot No.: TUV141013-01" sample met the overall migration requirements under Commission Regulation (EU) No 10/2011 – "Plastic materials and articles shall not transfer their constituents to foodstuffs in quantities exceeding 10 milligrams of total constituents released per dm<sup>2</sup> of food contact surface (mg/dm<sup>2</sup>) (overall migration limit)".

MS TAN SER LING  
TECHNICAL EXECUTIVE

DR XIAO ZHOU  
PRODUCT MANAGER  
MICROCONTAMINATION DIAGNOSIS  
CHEMICAL & MATERIALS

Cc: YTY Industry (Manjung)Sdn Bhd



TEST REPORT:


09 JAN 2014



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SUD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SUD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SUD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SUD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SUD PSB or to the report or results furnished by TÜV SUD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SUD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011



Product Certificates



SGMP Company Limited

บริษัท เอสจีเอ็มพี จำกัด

181 Moo 6, Tambol Kampaengphet, Amphur Rattaphum, Songkhla 90180 Thailand. Tel: (074) 834065-6, Fax: (074) 498595  
เลขที่ 181 หมู่ที่ 6 ตำบลกำแพงเพชร อำเภอรัตนบุรี จังหวัดสงขลา 90180 โทร. (074) 834065-6 แฟกซ์ (074) 498595

DECLARATION OF CONFORMITY

Product Name : Nitrile Examination Gloves (Non-Sterile)  
 Type : Ambidextrous  
 Powder-Free  
 Black, White & Blue

Manufacturer's Name : S G M P Company Limited  
 Manufacturer's Address : 181 Moo 6, Tambol Kampaengphet  
 Amphur Rattaphum, Songkhla 90180  
 Thailand

Document No. : SGMP-CE-DC-NITRILE-001  
 Classification : Class I

Brand : Sunrise Nitrile Examination Gloves  
 Sunrise Latex Examination Gloves

I, the undersign, hereby declare that the medical device(s) specified above conforms to the Essential Requirements listed in Annex VII of Medical Devices Directive (MDD) 93/42/ECC and bears the mark



This Declaration of Conformity is valid in connection with the release document for the respective batch produced devices. The above-mentioned declaration of conformity is under the responsibility of Manufacturer.

Danny Chiong  
 RAQA Officer

SGMP Company Limited  
 26 June 2020

Test Report No.: 60393368-001

Page 9 of 11

For use in contact with aqueous foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M002	
Distilled Water	Reflux temperature for 7 hours	0.1	0.7	20
	Succeeding 2 hours of extraction	0.1	0.2	1
<b>Comment</b>	--	--	Pass	--

For use in contact with fatty foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M002	
n-Hexane	Reflux temperature for 7 hours	0.1	0.4	175
	Succeeding 2 hours of extraction	0.1	n.d.	4
<b>Comment</b>	--	--	Pass	--

For use in contact with aqueous foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M003	
Distilled Water	Reflux temperature for 7 hours	0.1	0.7	20
	Succeeding 2 hours of extraction	0.1	n.d.	1
<b>Comment</b>	--	--	Pass	--

For use in contact with fatty foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M003	
n-Hexane	Reflux temperature for 7 hours	0.1	0.4	175
	Succeeding 2 hours of extraction	0.1	n.d.	4
<b>Comment</b>	--	--	Pass	--

Test Report No.: 60393368-001

Page 10 of 11

For use in contact with aqueous foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M004	
Distilled Water	Reflux temperature for 7 hours	0.1	0.8	20
	Succeeding 2 hours of extraction	0.1	n.d.	1
<b>Comment</b>	--	--	Pass	--

For use in contact with fatty foods:

Extractants	Test Condition	RL (mg/Inch <sup>2</sup> )	Result (mg/Inch <sup>2</sup> )	Permissible Limit (mg/Inch <sup>2</sup> )
			M004	
n-Hexane	Reflux temperature for 7 hours	0.1	0.4	175
	Succeeding 2 hours of extraction	0.1	n.d.	4
<b>Comment</b>	--	--	Pass	--

Abbreviation: n.d. denotes Not Detected (<RL)  
RL denotes Reporting Limit  
mg/Inch<sup>2</sup> denotes Milligram per square inch

Test Report No.: 60393368-001

Page 7 of 11

Test No.	T003		T004	
Material No.	M003		M004	
Tested no.	Force at break Before ageing (N)	Force at break After ageing (N)	Force at break Before ageing (N)	Force at break After ageing (N)
1	99.13	7.97	6.71	5.20
2	7.32	7.32	7.37	5.83
3	7.15	7060	6.57	5.49
4	8.16	7085	7.70	5.37
5	9.04	7072	7.16	5.57
6	8.59	7.84	6.40	5.34
7	7.86	7.33	7.70	4.48
8	8.33	8.29	5.73	4.68
9	7.74	7.39	7.05	5.85
10	9.65	9.11	7.58	4.45
11	7.89	9.26	7.27	5.33
12	7.84	6.87	6.35	4.58
13	9.59	7.73	8.26	5.73
Median result	8.33	7.87	7.06	5.22
Conclusion	Pass	Pass	Pass	Pass

Abbreviation: N denotes Newton

Remark:

1. Median values of force at break

Force at Break (newton)		
a)	b)	c)
≥ 9.0	≥ 6.0	≥ 3.0

a) Requirements for all surgical gloves

b) Requirements for all examination gloves, except gloves made from thermoplastic materials (e.g. Polyvinylchloride, Polyethylene)

c) Requirements for gloves made from thermoplastic materials (e.g. Polyvinylchloride, Polyethylene)

Test Report No.: 60393368-001

Page 8 of 11

**Proteins extraction**

Test method: With reference to EN 455-3: 2015 (E), Annex A.

Test result:

Test No.	Material No.	Test parameter	Unit	MDL	Test result
T001	M001	Proteins	µg/g	3	n.d.
T002	M002	Proteins	µg/g	3	n.d.
T003	M003	Proteins	µg/g	3	n.d.
T004	M004	Proteins	µg/g	3	n.d.

Abbreviation: µg/g denotes Microgram per gram

MDL denotes Method Detection Limit

Remark: The leachable protein level shall be "As Low As Reasonable Practicable" (ALARP)

**US FDA 21 CFR 177.2600 (Rubber Articles) – Determination of Amount of Extractives**

Test Method: With reference to US FDA 21 CFR 177.2600.

For use in contact with aqueous foods:

Extractants	Test Condition	RL (mg/inch <sup>2</sup> )	Result (mg/inch <sup>2</sup> )	Permissible Limit (mg/inch <sup>2</sup> )
			M001	
Distilled Water	Reflux temperature for 7 hours	0.1	0.5	20
	Succeeding 2 hours of extraction	0.1	n.d.	1
Comment	--	--	Pass	--

For use in contact with fatty foods:

Extractants	Test Condition	RL (mg/inch <sup>2</sup> )	Result (mg/inch <sup>2</sup> )	Permissible Limit (mg/inch <sup>2</sup> )
			M001	
n-Hexane	Reflux temperature for 7 hours	0.1	0.2	175
	Succeeding 2 hours of extraction	0.1	n.d.	4
Comment	--	--	Pass	--



Test Report No.: 60393368-001

Page 5 of 11

Test No.:	T003		T004	
Material No.:	M003		M004	
Tested no.	Length (mm)	Width (mm)	Length (mm)	Width (mm)
1	245	110	245	116
2	246	110	244	116
3	245	110	245	119
4	245	110	244	116
5	245	110	245	116
6	245	110	245	115
7	245	111	245	116
8	246	110	244	117
9	245	110	245	116
10	246	110	245	116
11	245	110	245	116
12	245	110	245	116
13	245	111	245	116
Median result	245	110	245	116
Conclusion	Pass	Pass	Pass	Pass

Abbreviation: mm denotes millimeter

Remark: Dimension limit reference to table

Size	Median length (mm)	Median width (mm)
Extra small	≥ 240	≤ 80
Small		80±10
Medium		95±10
Large		110±10
Extra large		≥ 110

Test Report No.: 60393368-001

Page 6 of 11

**Force at break**

Test method: EN 455-2: 2015

**Test results:**

Test No.	T001		T002	
Material No.	M001		M002	
Tested no.	Force at break Before ageing (N)	Force at break After ageing (N)	Force at break Before ageing (N)	Force at break After ageing (N)
1	6.31	6.33	10.11	10.98
2	7.22	6.81	9.45	8.67
3	6.47	5.92	9.75	10.57
4	7.46	6.72	8.33	9.41
5	5.56	5.35	10.80	8.94
6	5.47	5.93	8.48	10.75
7	6.18	6.70	8.80	9.46
8	6.32	4.80	10.00	8.65
9	5.72	6.21	9.95	9.83
10	6.71	5.87	9.12	9.58
11	6.45	6.68	9.11	8.90
12	5.87	5.95	11.02	8.40
13	6.53	6.52	8.97	10.54
Median result	6.33	6.14	9.53	9.59
Conclusion	Pass	Pass	Pass	Pass

Test Report No.: 60393368-001

Page 3 of 11

**Freedom from holes**

Test method: With reference to EN 455-1: 2000

Test result:

Gloves Size	Tested samples	No. of samples for Non-compliance	Conclusion
S	200 pcs.	1	Pass
M	200 pcs.	0	Pass
L	200 pcs.	0	Pass
XL	200 pcs.	0	Pass

Remark:

- All samples were selected and supplied by the client.
- The batch size of the gloves supplied was not stated by the client. In accordance with BS EN 455-1, a batch size between 35,001 to 150,000 was chosen, and therefore 50 gloves per stage were tested for perforations using General Inspection Level I at an AQL of 1.5% with reference to table;

Stage No.	Cumulative no. tested	Accept	Reject
First	50	0	4
Second	100	1	6
Third	150	3	8
Fourth	200	5	9
Fifth	250	9	19

Test Report No.: 60393368-001

Page 4 of 11

**Dimension test**

Test method: EN 455-2: 2015

Test results:

Test No.:	T001		T002	
Material No.:	M001		M002	
Tested no.	Length (mm)	Width (mm)	Length (mm)	Width (mm)
1	246	85	244	94
2	246	86	244	94
3	245	85	246	94
4	246	85	246	94
5	245	85	245	94
6	245	85	244	93
7	245	84	245	94
8	245	85	245	94
9	245	85	244	94
10	244	82	244	94
11	245	85	244	94
12	243	85	244	95
13	244	85	244	94
Median result	245	85	245	94
Conclusion	Pass	Pass	Pass	Pass





Test Report No. 4605024

Date : 23-May-2020

Page 3 of 3

SAMPLE/ATTACHMENT PICTURE



\*\*\*\*\* End of Report \*\*\*\*\*

"Any holder of this document is advised that should client or third party information be supplied with respect to the goods or sample, SGS may, at its discretion, attached or indicate such information to the report but SGS makes no warranties or accepts no liable for the veracity or lack thereof of such information."

4306737

"This document is issued by the Company subject to its General Conditions of Service printed separately, available on request or accessible at <http://www.sgs.com/Forms-and-Conditions/Forms-and-Conditions.aspx> and its electronic form attached herewith, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/Forms-and-Conditions/Forms-and-Conditions.aspx>

Attention is drawn to the limitation of liability, indemnification and jurisdiction clause defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its inspection only and within the limits of Client's instructions. It also, the Company's sole responsibility is to its Client and this document shall not constitute evidence in a court of law. Any unauthorized alteration, reuse or modification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."

Unless otherwise stated the results shown in this report refer only to the samples tested and such samples are retained for 30 days only.

SGS (Thailand) Limited Laboratory Services 41/23 Soi Rama II 50 Rama II Road Chongnonsi Yonhwa Bangkok 10120  
 T +66 (0)2 583 35 41 F +66 (0)2 294 74 84 883 07 50 www.sgs.com

Member of the SGS Group

Products



TÜVRheinland®  
Precisely Right.

Test Report No.: 60393368-001

Page 2 of 11

**Sampling Information:**

Inspection Method: No inspection, submitted sample by client  
 Inspection level: N/A  
 AQL: N/A  
 Sample size: N/A

**Material list:**

Material No.	Material	Color	Location
M001	Nitrile Glove	Purple	Size S
M002	Nitrile Glove	Blue	Size M
M003	Nitrile Glove	Blue	Size L
M004	Nitrile Glove	Blue	Size XL

2020/5/10 510(k) Premarket Notification



**510(k) Premarket Notification**

[FDA Home](#)
[Medical Devices](#)
[Databases](#)

[510\(k\)](#)
[DeNovo](#)
[Registration & Listing](#)
[Adverse Events](#)
[Recalls](#)
[PMA](#)
[HDE](#)
[Classification](#)
[Standards](#)

[CFR Title 21](#)
[Radiation-Emitting Products](#)
[X-Ray Assembler](#)
[Medsun Reports](#)
[CLIA](#)
[TPLC](#)

[New Search](#)
[Back To Search Results](#)

**Device Classification Name** [Polymer Patient Examination Glove](#)<sup>22</sup>  
**510(K) Number** K000988  
**Device Name** NON-STERILE POWDER-FREE, BLUE NITRILE EXAMINATION GLOVES  
**Applicant** SGMP CO., LTD.  
 181 MOO 8, TAMBOL  
 KAMPAENGPETCH, RATTAPHUM  
 Songkhla, TH 90180  
**Applicant Contact** Cheah Chor Hee  
**Correspondent** SGMP CO., LTD.  
 181 MOO 8, TAMBOL  
 KAMPAENGPETCH, RATTAPHUM  
 Songkhla, TH 90180  
**Correspondent Contact** Cheah Chor Hee  
**Regulation Number** [880.6250](#)<sup>23</sup>  
**Classification Product Code** [LZA](#)<sup>24</sup>  
**Date Received** 03/17/2000  
**Decision Date** 04/20/2000  
**Decision** Substantially Equivalent (SESE)  
**Regulation Medical Specialty** General Hospital  
**510k Review Panel** General Hospital  
**Type** Traditional  
**Reviewed By Third Party** No  
**Combination Product** No

Links on this page:

- <http://www.addthis.com/bookmark.php?v=152&username=fdamain>
- <http://www.addthis.com/bookmark.php>
- <https://www.fda.gov/>
- <https://www.fda.gov/Medical-Devices>
- <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
- </scripts/cdrh/devicesatfda/index.cfm>
- </scripts/cdrh/cfdocs/cfPMN/pmn.cfm>
- </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
- </scripts/cdrh/cfdocs/cfRL/rl.cfm>
- </scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm>
- </scripts/cdrh/cfdocs/cfRES/res.cfm>
- </scripts/cdrh/cfdocs/cfPMA/pma.cfm>
- </scripts/cdrh/cfdocs/cfHDE/hde.cfm>
- </scripts/cdrh/cfdocs/cfPCD/classification.cfm>

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K000988>



Test Report No. 4605024

Date : 23-May-2020

Page 1 of 3

Client :

The following sample(s) was/were submitted and identified by client as:

Sample Name : Disposable Rubber Glove  
 Color : White (Nature)  
 Manufacturer/Vendor : SAS Rubber Product Co.,Ltd.  
 Country of Origin : Thailand  
 Country of Destination : Around the world

The following sample(s) was/were identified by SGS as:

SGS Sample No. : 4778104  
 Sample Condition : Sample is contained in a plastic bag.  
 Quantity Submitted : 1 pair

Sample Receiving Date : 19-May-2020  
 Testing Period : 19-May-2020 to 23-May-2020

Test Method & Results : Please refer to next page(s).

### Test Requested & Result Summary

Test Requested : Please refer to the result summary (Test parameter(s) was/were selected by client).

Result Summary:

Test Requested	Conclusion
US FDA 21 CFR 177.2600 (Rubber Articles)- Determination of Amount of Extractives	-
- Distilled Water Extractants	PASS
- n-Hexane Extractants	PASS

Remark: Test results in this report are applicable for the item tested and reflects the tested sample as received.

Signed for and on behalf of  
SGS (Thailand) Limited

Rutchaporn Mungsom  
Laboratory manager - Toy and Hardgood

"Any holder of this document is advised that should client or third party information be supplied with respect to the goods or sample, SGS may, at its discretion, attach or indicate such information to the report but SGS makes no warranties or accepts no liability for the veracity or lack thereof of such information."

This document is issued by the Company subject to its General Conditions of Service printed separately, available on request or accessible at <http://www.sgs.com/Service/Forms-and-Conditions.aspx> and its electronic forms documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/Service/Forms-and-Conditions/Forms-and-Conditions.aspx>

Attention is drawn to the limitation of liability, indemnification and jurisdiction clause defined therein. Any holder of this document is advised that the information contained therein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not constitute a recommendation, endorsement or any other form of advice. This document cannot be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, or by any information storage and retrieval system, without the prior written permission of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and prohibited by law. Please refer to the General Conditions of Service.

Unless otherwise stated the results shown in this test report refer only to the samples, tested and such samples are retained for 30 days only.

SGS (Thailand) Limited Laboratory Services 41/02 Soi Rama 91 58 Rama 9 Road Chongnonsi Yankwut Bangkok 10120  
 T +66 (0)2 563 05 41 F +66 (0)2 564 74 85-80 E +66 (0)2 564 74 84 683 87 58 www.sgs.com

Member of the SGS Group

**Indications for Use**

510(k) Number (if known): K042879

Device Name: NON-STERILE, POWDER-FREE BLUE NITRILE EXAMINATION GLOVES WITH VANILLA SCENTING

Indications For Use: **This is a medical glove to be worn on the hand of health care and similar personnel to prevent contamination between health care personnel and the patient.**

Prescription Use \_\_\_\_\_ AND/OR Over-The-Counter Use X  
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

*[Signature]*  
(Division Sign-Off)  
 Division of Anesthesiology, General Hospital,  
 Injection Control, Dental Devices

510(k) Number: K042879

2020/8/10 510(k) Premarket Notification

**FDA** U.S. FOOD & DRUG ADMINISTRATION

[FDA Home](#) [Medical Devices Database](#)

**510(k) Premarket Notification**

[510\(k\) De Novo Registration & Listing](#) [Adverse Events](#) [Recalls](#) [PMA](#) [HDE](#) [Classification](#) [Standards](#)  
[CFR Title 21](#) [Radiation-Emitting Products](#) [X-Ray Assembler](#) [Medsun Reports](#) [CLIA](#) [TPLC](#)

[New Search](#) [Back To Search Results](#)

Device Classification Name: [Polymer Patient Examination Glove](#)<sup>22</sup>

510(k) Number: K020317

Device Name: NON-STERILE POWDER FREE GREEN BARRIER-PRO SYNTHETIC BUTADIENE COPOLYMER EXAMINATION GLOVE, COLOR: BLUE

Applicant: SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434

Applicant Contact: Janna Tucker

Correspondent: SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434

Correspondent Contact: Janna Tucker

Regulation Number: [880.6250](#)<sup>23</sup>

Classification: [LZA](#)<sup>24</sup>

Product Code

Date Received: 01/30/2002

Decision Date: 02/14/2002

Decision: Substantially Equivalent (SESE)

Regulation Medical Specialty: General Hospital

510k Review Panel: General Hospital

Summary: [Summary](#)<sup>25</sup>

Type: Traditional

Reviewed By Third Party: No

Combination Product: No

---

Links on this page:

- <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
- <http://www.addthis.com/bookmark.php>
- <https://www.fda.gov/>
- <https://www.fda.gov/Medical-Devices>
- <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
- </scripts/cdrh/devicesatfda/index.cfm>
- </scripts/cdrh/cfdocs/cfPMN/pmnm.cfm>
- </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
- </scripts/cdrh/cfdocs/cfRL/rl.cfm>

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnm.cfm?ID=K020317>

1/3





DEPARTMENT OF HEALTH &amp; HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

JAN 13 2005

SGMP Company Limited  
C/O Ms. Janna P. Tucker  
Official Correspondent  
Tucker & Associates  
198 Avenue De La D' Emerald  
Sparks, Nevada 89434-9550

Re: K042879  
Trade/Device Name: Non-Sterile, Powder-Free Blue Nitrile Examination  
Gloves with Vanilla Scenting  
Regulation Number: 880.6250  
Regulation Name: Patient Examination Glove  
Regulatory Class: I  
Product Code: LZA  
Dated: December 31, 2004  
Received: January 4, 2005

Dear Ms. Tucker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Ms. Tucker

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

Chiu Lin, Ph.D.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

Page 2 – Ms. Tucker

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Chiu Lin, Ph.D.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

INDICATION FOR USE STATEMENT

Applicant : SGMP Company Limited

510K NUMBER: K072400


Device Name : Non-Sterile, Powder Free Black Nitrile Examination Gloves

Indication For Use :

The Non-sterile Powder Free Black Nitrile Examination Gloves , is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Prescription Use ..... AND / OR Over-The-Counter.....  
(Part 21 CFR 801.Subpart D) 21 CFR 801 Subpart C

Concurrence of CDRH , Office of Device Evaluation (ODE)

  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: K072400

## Product Certificates

2020/8/10 510(k) Premarket Notification

**FDA** U.S. Food & Drug Administration

[FDA Home](#) [Medical Devices](#) [Databases](#)

**510(k) Premarket Notification**

[\(510\(k\)\)](#) [De Novo](#) [Registration & Listing](#) [Adverse Events](#) [Recalls](#) [PMA](#) [HDE](#) [Classification](#) [Standards](#)

[Super Search](#) [CDRH](#) [CFR Title 21](#) [Radiation-Emitting Products](#) [X-Ray Assembler](#) [Medsun Reports](#) [CLIA](#) [TPLC](#)

[New Search](#) [Back To Search Results](#)

Device Classification Name	<a href="#">Polymer Patient Examination Glove</a> <sup>22</sup>
510(K) Number	K042879
Device Name	NON-STERILE POWDER-FREE BLUE NITRILE EXAMINATION GLOVES, WITH VANILLA SCENTING
Applicant	SGMP CO., LTD. 198 AVENUE DE LA D'EMERALD Sparks, NV 89434
Applicant Contact Correspondent	Janna P Tucker SGMP CO., LTD. 198 AVENUE DE LA D'EMERALD Sparks, NV 89434
Correspondent Contact	Janna P Tucker
Regulation Number	<a href="#">880.6250</a> <sup>23</sup>
Classification Product Code	<a href="#">LZA</a> <sup>24</sup>
Date Received	10/18/2004
Decision Date	01/13/2005
Decision	Substantially Equivalent (SESE)
Regulation Medical Specialty	General Hospital
510k Review Panel	General Hospital
Summary	<a href="#">Summary</a> <sup>25</sup>
Type	Traditional
Reviewed By Third Party	No
Combination Product	No

Links on this page:

- <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
- <http://www.addthis.com/bookmark.php>
- <https://www.fda.gov/>
- <https://www.fda.gov/Medical-Devices>
- <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
- </scripts/cdrh/devicesatfda/index.cfm>
- </scripts/cdrh/cfdocs/cfPMN/pmn.cfm>
- </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
- </scripts/cdrh/cfdocs/cfRL/rl.cfm>
- </scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm>
- </scripts/cdrh/cfdocs/cfRES/res.cfm>
- </scripts/cdrh/cfdocs/cfPMA/pma.cfm>

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K042879> 1/3

## Product Certificates

K042879

JAN 13 2005

**APPENDIX L**

**510(k) SUMMARY**  
SUMMARY OF THE SAFETY AND EFFECTIVENESS FOR POWDER-FREE BLUE NITRILE EXAMINATION WITH VANILLA SCENTING

**Submitted For :** SGMP Company Limited, 181 Moo 6, Tambol Kampaengpetch, Rattaphum, Songkhla 90180, Thailand.

**Submitted By:** Tucker & Associates  
Official Correspondent for SGMP Co Ltd  
Janna P. Tucker, President – CEO  
198 Avenue de la D'emerald, Sparks, NV 89434-9550  
Phone No : 775-342-2612 Fax No : 775-342-2613  
E-mail: Tuckerjan@aol.com

**Equivalent Predicate Device:** POWDER FREE NITRILE EXAM GLOVES which was granted a 510 (k) # K000868 as shown in APPENDIX M

This summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990.

**Device Information:**  
Trade Name – NON-STERILE POWDER FREE BLUE NITRILE EXAMINATION GLOVES WITH VANILLA SCENTING  
Common Name - Exam gloves  
Classification Name - Patient examination glove (per 21 CFR 880.6250)  
Classification Information - Class I nitrile patient examination glove 80LZA, powder free and meeting all the requirements of ASTM-D6319-00aE1 Standard Specification for Nitrile Examination Gloves for Medical Application.

**Device Description:**  
Class I nitrile patient examination gloves 80LZA, powder free and meeting all the requirements of ASTM-D6319-00aE1 Standard Specification for Nitrile Examination Gloves for Medical Application.

**Intended Use of Device:**  
A medical glove to be worn on the hand of the health care and similar personnel to prevent contamination between health care personnel and patient.



K072400

**Conclusion:**

The data presented indicate the Non-sterile Powder-Free Black Nitrile Examination Glove meets the following Standards:

1. ASTM D6319-00aE3, Standard Specification For Nitrile Gloves.
2. ISO 2859-1, Standard for Water Leak Test and/or ASTM D5151-06, Standard Test Method for Detection of Holes in Medical Gloves.
3. ASTM D6124-06, Standard Test Method for Residual Powder on Medical Gloves.
4. Biocompatibility Testing on White rabbits and Guinea Pigs.
5. Labeling meets FDA Specifications
6. Except for Color, this glove is equivalent to K000868.

APPENDIX K

Page 2 of 2



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

MAR 27 2008

SGMP Company, Limited  
C/O Ms. Janna P. Tucker  
Official Correspondent/ United States Agent  
Tucker & Associates  
198 Avenue De La D' Emerald  
Sparks, Nevada 89434-9550

Re: K072400  
Trade/Device Name: Non-Sterile, Powder Free Black Nitrile Examination Gloves  
Regulation Number: 880.6250  
Regulation Name: Patient Examination Glove  
Regulatory Class: I  
Product Code: LZA  
Dated: March 11, 2008  
Received: March 13, 2008

Dear Ms. Tucker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).

2020/8/10 510(k) Premarket Notification



[FDA Home](#)<sup>3</sup> [Medical Devices](#)<sup>4</sup> [Database](#)<sup>5</sup>

**510(k) Premarket Notification**

[610\(k\)7](#) [DeNovo8](#) [Registration & Listing9](#) [Adverse Events10](#) [Recalls11](#) [PMA12](#) [HDE13](#) [Classification14](#) [Standards15](#)  
[CFR Title 2116](#) [Radiation-Emitting Products17](#) [X-Ray Assembler18](#) [Medsun Reports19](#) [CLIA20](#) [TPLC21](#)

[New Search](#) [Back To Search Results](#)

**Device Classification Name** [Polymer Patient Examination Glove22](#)  
**510(K) Number** K072400  
**Device Name** NON-STERILE, POWDER-FREE, BLACK NITRILE EXAMINATION GLOVES  
**Applicant** SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434  
**Applicant Contact** Janna P Tucker  
**Correspondent** SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434  
**Correspondent Contact** Janna P Tucker  
**Regulation Number** [880.625023](#)  
**Classification Product Code** [LZA24](#)  
**Date Received** 08/27/2007  
**Decision Date** 03/27/2008  
**Decision** Substantially Equivalent (SESE)  
**Regulation Medical Specialty** General Hospital  
**510k Review Panel** General Hospital  
**Summary** [Summary25](#)  
**Type** Traditional  
**Reviewed By Third Party** No  
**Combination Product** No

**Links on this page:**

- <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
- <http://www.addthis.com/bookmark.php>
- <https://www.fda.gov/>
- <https://www.fda.gov/Medical-Devices>
- <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
- </scripts/cdrh/devicesatfda/index.cfm>
- </scripts/cdrh/cfdocs/cfPMN/pmn.cfm>
- </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
- </scripts/cdrh/cfdocs/cfRL/r1.cfm>
- </scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm>
- </scripts/cdrh/cfdocs/cfRES/res.cfm>
- </scripts/cdrh/cfdocs/cfPMA/pma.cfm>
- </scripts/cdrh/cfdocs/cfHDE/hde.cfm>
- </scripts/cdrh/cfdocs/cfPCD/classification.cfm>

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K072400> 1/3

MAR 27 2008 APPENDIX K

**510(k) SUMMARY** *K072400*

**SUMMARY OF THE SAFETY AND EFFECTIVENESS FOR NON-STERILE, POWDER FREE BLACK NITRILE EXAMINATION GLOVES**

**Submitted For:** SGMP Company Limited, 181 Moo 6, Tambol Kampaengpetch, Rattaphum, Songkhla 90180, Thailand

**Submitted By:** Tucker & Associates  
 Official Correspondent for SGMP Co Ltd  
 Janna P. Tucker, President - CEO  
 198 Avenue de la D'emerald, Sparks, NV 89434-9550  
 Phone No : 775-342-2612 Fax No : 775-342-2613  
 E-mail: Tuckerjan@aol.com

**Equivalent Predicate Device:** POWDER FREE BLUE NITRILE EXAM GLOVES which was granted a 510(k) # K000868

This summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990.

**Device Information:**  
 Trade Name - Non-Sterile, Powder Free Black Nitrile Examination Gloves

Common Name - Exam gloves  
 Classification Name - Patient examination glove (per 21 CFR 880.6250)  
 Classification Information - Class I Nitrile examination glove 80LZA, powder free and meeting all the requirements of ASTM D6319-00aE3 Standard Specification for Nitrile Examination Gloves for Medical Application.

**Device Description:**  
 Class I Nitrile examination gloves 80LZA, powder free and meeting all the requirements of ASTM D6319-00aE3 Standard Specification for Nitrile Examination Gloves for Medical Application.

**Intended Use of Device:**  
 A medical glove to be worn on the hand of the health care and similar personnel to prevent contamination between health care personnel and patient.

*Page 1 of 2*

## Product Certificates

K101822

**CONCLUSION:**


The data presented indicates the Powder-Free, Blue Nitrile Examination Gloves, Tested for Use With Chemotherapy Drugs Labeling Claim, (Non-Sterile), K101822 is equivalent to K082957, Non-Sterile, Powder-free Nitrile Examination Gloves, Blue with Polymer Coating, Tested for use with Chemotherapy Drugs.

**It should be noted that testing for use with two chemotherapy drugs had extremely low permeation times as follows: CARMUSTINE (BCNU) @ 0.49 minutes, and THIOTEPA @ 2.61 minutes. Therefore, these gloves are not approved for use when using those chemo drugs.**

These gloves do meet the following recognized standards unless otherwise noted:

- ASTM D6319-00a00A(2005), Standard Specification for Nitrile Gloves
- ISO 2859-1, Standard for Water Leak Test and/or ASTM D5151-06, Test Method for Detection of Holes in Medical Gloves
- ASTM D6124-06, Standard Test Method for Residual Powder on Medical Gloves.
- Biocompatibility Testing on White Rabbits and Guinea Pigs
- Labeling meets FDA requirement
- Substantially equivalent to Siam Sempermed Corp. Ltd. K082957
- ASTM D6978-05 Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs.

## Product Certificates

 **DEPARTMENT OF HEALTH & HUMAN SERVICES** Public Health Service

---

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -W066-G609  
Silver Spring, MD 20993-0002

NOV 19 2010

SGMP Company, Limited  
C/O Ms. Janna P. Tucker  
Tucker & Associates  
198 Avenue De La D'Emerald  
Sparks, Nevada 89434

Re: K101822  
Trade/Device Name: Non-Sterile, Powder Free Blue Nitrile Examination Gloves  
Tested for use with Chemotherapy Drugs  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Patient Examination Glove  
Regulatory Class: I  
Product Code: LZA/LZC  
Dated: October 22, 2010  
Received: October 21, 2010

Dear Ms. Tucker:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the [Federal Register](#).



K101822

8. Substantial Equivalence Discussion :

Characteristic and Parameter	SGMP Company Limited Non-sterile, Powder-free Blue Nitrile Examination Gloves. Tested for use with Chemotherapy Drugs	Siam Sempemed Corp.Ltd. Non-sterile, Powder-free Nitrile Examination Glove, Blue with Polymer Coating, Tested for use with Chemotherapy Drugs 510 K # K082957	Substantial Equivalence (SE)
Devise Class	I	I	SE
Product Code	LZA	LZA	SE
Glove Color	Blue	Blue	SE
Dimensions	Meets ASTM D6319-00a-05	Meets ASTM D6319-00a-05	SE
Physical Properties	Meets ASTM D6319-00a-05	Meets ASTM D6319-00a-05	SE
Freedom From Pinholes	Meets ASTM D6319-00a-05	Meets ASTM D6319-00a-05	SE
Powder-free Residue	Meets ASTM D6124-06	Meets ASTM D6124-06	SE

K101822

Biocompatibility Test	Passes Primary Skin Irritation in Rabbits	Passes Primary Skin Irritation in Rabbits	SE
	Passes Guinea Pig Maximization	Passes Guinea Pig Sensitization	SE
Chemotherapy Drugs Tests	Meets ASTM D6978-05 Cisplatin > 240 mins Cyclophosphamide > 240 mins Doxorubicin Hydrochloride > 240 mins Etoposide > 240 mins Flurouracil > 240 mins Paclitaxel > 240 mins Vencristine Sulfate > 240 mins Dacarbazine > 240 mins Methotrexate > 240 mins	Meets ASTM D6978-05 Cisplatin > 240 mins Cyclophosphamide > 240 mins Doxorubicin Hydrochloride > 240 mins Etoposide > 240 mins Flurouracil > 240 mins Paclitaxel > 240 mins Vencristine Sulfate > 240 mins Dacarbazine > 240 mins Methotrexate > 240 mins	SE

K101822

510 (k) SUMMARY *APPENDIX M*  
 (As Required by 21 section 807.92 (c))

NOV 19 2010

1. **Submitted For :**  
 SGMP Company Limited  
 181 Moo 6, Tambol Kampaengphet,  
 Rattaphum, Songkhla 90180 Thailand,
2. **Submitted By :**  
 Tucker & Associates  
 Official Correspondent for SGMP Company Limited  
 Janna P. Tucker, President - CEO  
 198, Avenue de la D'Emerald, Sparks,  
 NV 89434-9550  
 Phone No : 775-342-2612  
 Fax No : 775-342-2613  
 Email : [Tuckerjan@aol.com](mailto:Tuckerjan@aol.com)
3. **Device Trade or Proprietary Name :**  
 Non-sterile, Powder-free Blue Nitrile Examination Gloves, Tested for use with  
 Chemotherapy Drugs.
4. **Device Common Name :**  
 Examination Gloves
5. **Device Classification Name :**  
 Patient Examination Gloves (per 21CFR 880.6250)
6. **Device Description :**  
 Non-sterile, Powder-free Blue Nitrile Examination Gloves, Tested for use with  
 Chemotherapy Drugs.
7. **Intended Use of Device :**  
 A disposable medical glove to be worn on the hand of the healthcare and similar  
 personnel to prevent contamination between healthcare personnel and patient.

*ATCH 2  
 APPENDIX M revised 10-12-10, Janna Tucker  
 Now shows Equivalence to K082957*

K101822

The Non-sterile Powder Free Blue Nitrile Examination Gloves Tested for use with  
 Chemotherapy Drugs, is a disposable device intended for medical purposes that is worn on the  
 examiner's hand or finger to prevent contamination between patient and examiner.

Chemotherapy Drug Permeation (Breakthrough Detection Time) in Minutes  
 The following chemical have been tested with these gloves.

Test Chemotherapy Drug	Average Breakthrough Detection Time (Minutes)
<b>*Carmustine (BCNU)</b>	<b>0.49 minutes</b>
Cisplatin	>240 minutes
Cyclophosphamide (Cytoxan)	>240 minutes
Doxorubicin Hydrochloride	>240 minutes
Etoposide (Toposar)	>240 minutes
Flurouracil	>240 minutes
Paclitaxel (Taxol)	>240 minutes
<b>*Thiotepa</b>	<b>2.61 minutes</b>
Vincristine Sulfate	>240 minutes
Dacarbazine (DTIC)	>240 minutes
Methotrexate	>240 minutes

\*Please note that Carmustine (BCNU) and Thiotepa have extremely low  
 permeation times of 0.49 and 2.61 minute only.

# Product Certificates

U.S. Department of Health & Human Services

Follow FDA | En Español

**U.S. FOOD & DRUG ADMINISTRATION**

Home | Food | Drugs | Medical Devices | Radiation-Emitting Products | Vaccines, Blood & Biologics | Animal & Veterinary | Cosmetics | Tobacco Products

### Establishment Registration & Device Listing

1 result found for Owner Operator Name : sgmp

Establishment Name	Registration Number	Current Registration Yr
SGMP CO., LTD. THAILAND	9515905	2020

- Polymer Patient Examination Glove - Black Nitrile
- Latex Patient Examination Glove
- Latex Patient Examination Glove - Blue Bubble Gum
- Latex Patient Examination Glove
- Latex Patient Examination Glove

Show all 22 Listings For SGMP CO., LTD. THAILAND

Can't find what you're looking for? [Try a new search](#)

Page Last Updated: 09/24/2020  
 Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.  
 Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | العربية | English

U.S. Food and Drug Administration  
 10903 New Hampshire Avenue  
 Silver Spring, MD 20993  
 Ph. 1-888-INFO-FDA (1-888-463-6332)  
 Contact FDA

U.S. Department of Health & Human Services

# Product Certificates

2020/8/10 510(K) Premarket Notification

**FDA** U.S. Food & Drug Administration

[FDA Home](#) [Medical Devices](#) [Database](#)

### 510(k) Premarket Notification

[\(510\(k\)\)](#) [De Novo](#) [Registration & Listing](#) [Adverse Events](#) [Recalls](#) [PMA](#) [HDE](#) [Classification](#) [Standards](#) [CFR Title 21](#) [Radiation-Emitting Products](#) [X-Ray Assembler](#) [Medsun Reports](#) [CLIA](#) [TFLC](#)

[New Search](#) [Back To Search Results](#)

**Device**  
**Classification Name** [Patient Examination Glove, Specialty](#)<sup>22</sup>  
**510(K) Number** K101822  
**Device Name** NON-STERILE POWDER-FREE BLUE NITRILE EXAMINATION GLOVES TESTED FOR USE WITH CHEMOTHERAPY DRUGS

**Applicant**  
 SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434

**Applicant Contact Correspondent**  
 Janna P Tucker  
 SGMP CO., LTD.  
 198 AVENUE DE LA D'EMERALD  
 Sparks, NV 89434

**Correspondent Contact**  
 Janna P Tucker

**Regulation Number** [880.6250](#)<sup>23</sup>  
**Classification** [L2C](#)<sup>24</sup>  
**Product Code**  
**Subsequent Product Code** [L2A](#)<sup>25</sup>

**Date Received** 08/30/2010  
**Decision Date** 11/19/2010  
**Decision** Substantially Equivalent (SESE)  
**Regulation Medical Specialty** General Hospital  
**510k Review Panel** General Hospital  
**Summary** [Summary](#)<sup>26</sup>  
**Type** Traditional  
**Reviewed By Third Party** No  
**Combination Product** No

---

**Links on this page:**

- <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdamain>
- <http://www.addthis.com/bookmark.php>
- <https://www.fda.gov/>
- <https://www.fda.gov/Medical-Devices>
- <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
- </scripts/cdrh/devicesatfda/index.cfm>
- </scripts/cdrh/cfdocs/cfPMN/pmn.cfm>
- </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
- </scripts/cdrh/cfdocs/cfRL/rl.cfm>

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K101822>

1/3



ZERTIFIKAT • CERTIFICATE • 認證證書 • CERTIFICADO • CERTIFICAT

UCB,F\_12.04 2012-02

# CERTIFICATE

No. QS5 18 06 52111 005



**Certificate Holder:** SGMP Company Limited  
181 Moo 6, Tambol Kampaengphet  
Amphur Rattaphum, Songkhla Pro. 90180  
THAILAND

**Certification Mark:**



**Scope of Certificate:** Design and Development, Production and Distribution of Non-sterile Powdered and Powder Free Examination Gloves

**Standard(s):** ISO 9001:2015

The Certification Body of TÜV SÜD America Inc. certifies that the company mentioned above has established and is maintaining a quality management system that meets the requirements of the listed standards.

**Report No.:** M2930

**Effective Date:** 2018-06-13  
**Expiry Date:** 2021-06-01



*Manuel Bradaric*  
Manuel Bradaric  
MHS Certification Manager

Page 1 of 1

TÜV SÜD America Inc.  
10 Centennial Drive  
Peabody, MA 01960  
USA



ZERTIFIKAT • CERTIFICATE • 認證證書 • CERTIFICADO • CERTIFICAT

UCB,F\_12.02 2012-02

# CERTIFICATE

No. QS6 18 06 52111 004



**Certificate Holder:** SGMP Company Limited  
181 Moo 6, Tambol Kampaengphet  
Amphur Rattaphum, Songkhla Pro. 90180  
THAILAND

**Certification Mark:**



**Scope of Certificate:** Design and Development, Production and Distribution of Non-sterile Powdered and Powder Free Examination Gloves


**Standard(s):** ISO 13485:2016

**Regulatory Authority:** Health Canada.  
See attached for listing of specific regulatory requirements.

The Certification Body of TÜV SÜD America Inc. certifies that the quality management system of the manufacturer listed above has been audited against the stated criteria and found to conform to those criteria for the scope of certification listed. Validity of this certificate can be obtained by visiting the website  
<http://www.tuv-sud-america.com/us-en/resource-center/customer-support/certificate-finder>

TÜV SÜD America Inc. is an MDSAP Recognized Auditing Organization.

**DUNS No:** 66-072-9989  
**Effective Date:** 2018-06-02  
**Expiry Date:** 2021-06-01



*Manuel Bradaric*  
Manuel Bradaric  
MHS Certification Manager

Page 1 of 2

TÜV SÜD AMERICA INC • 10 Centennial Drive • Peabody, MA 01960 USA • www.TUVamerica.com



ZERTIFIKAT • CERTIFICATE • 認證證書 • CERTIFICADO • СЕРТИФИКАТ • CERTIFICAT



## CERTIFICATE

No. QS6 18 06 52111 004

### Audit/Certification Criteria

#### Canada

- Medical Device Regulations SOR/98-282, Part 1

Effective Date: 2018-05-02  
 Expiry Date: 2021-05-01

Manuel Bradaric  
 Certification Manager MHS

Page 2 of 2

TÜV SÜD AMERICA INC • 10 Centennial Drive • Peabody, MA 01960 USA • www.TUVamerica.com



UCBLF\_12.02\_2012.02

Issued to: GGMP Company Limited  
 181 Moo6  
 Tambol Kampaenghet  
 Amphur Rattaphum  
 Songkhla 90180  
 Thailand

Notified Body: 2777
SATRA customer number: P1643

EU Type-Examination Certificate

**Certificate number: 2777/14166-01/E00-00**

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:  
 Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

<b>Product reference:</b>	<b>Description:</b>		
PF NBR 8mil	Powder free Nitrile Glove, 8mil with Pattern texture available in Black, Orange, Green and Yellow colours.		
<b>Sizes:</b>	<b>Classification:</b>		
Medium – 5558PF-M	EN ISO 374-1:2016/ Type C	Level	EN 374-4: 2013 Degradation %
Large – 5558PF-L	n-Heptane (J)	2	27
X-Large – 5558PF-XL	40% Sodium Hydroxide (K)	0	-58.7
XX-Large – 5558PF-XXL	90% Sulphuric Acid (L)	0	100.0
	<b>EN ISO 374-5: 2016</b>	<b>Result</b>	
	Protection against bacteria and fungi	Pass	
	Protection against viruses	N/A	

Standards/Technical specifications applied:  
 EN 420: 2003+A1: 2009; EN 388:2016+A1:2018; EN ISO 374-1:2016+A1:2018

Technical reports/Approval documents:  
 SATRA: SPC0247447/1626/2, CHM0247460/1626/EN/B, CHM0260647/1731/EN, CHM0262288/1738/SMcD/B, SPC0262693/1740, CHM0262288/1738/SMcD/A

Signed on behalf of SATRA:

Date first issued: 26/02/2020  
 Date of issue: 26/02/2020

Besjana Pillinci

Quincey Brown

Expiry date: 26/02/2025

Page 1 of 2

